

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069208.D
 Acq On : 27 Oct 2021 1:35
 Operator : JC/MD
 Sample : M4337-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DDC-2-PS

Quant Time: Oct 27 09:26:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	318970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	553523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	538105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	193975	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	210868	50.141	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.280%	
35) Dibromofluoromethane	8.024	113	163005	50.616	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.240%	
50) Toluene-d8	10.443	98	694872	53.927	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.860%	
62) 4-Bromofluorobenzene	12.734	95	246587	50.722	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	5601	3.569	ug/l	93
64) Tetrachloroethene	10.980	164	22980	6.629	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069208.D
Acq On : 27 Oct 2021 1:35
Operator : JC/MD
Sample : M4337-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DDC-2-PS

Quant Time: Oct 27 09:26:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

