

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069221.D
 Acq On : 27 Oct 2021 16:22
 Operator : JC/MD
 Sample : M4338-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5S

Quant Time: Oct 28 05:21:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	335018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	583126	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	563021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	202308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	219603	49.717	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.440%
35) Dibromofluoromethane	8.027	113	170881	50.368	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.740%
50) Toluene-d8	10.443	98	723007	53.262	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.520%
62) 4-Bromofluorobenzene	12.733	95	252984	49.396	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.800%
Target Compounds						
44) Trichloroethene	9.212	130	1338	0.352	ug/l	Qvalue 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069221.D
Acq On : 27 Oct 2021 16:22
Operator : JC/MD
Sample : M4338-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5S

Quant Time: Oct 28 05:21:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

