

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069222.D
 Acq On : 27 Oct 2021 16:47
 Operator : JC/MD
 Sample : M4338-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5D

Quant Time: Oct 28 05:21:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	342164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	598583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	581575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	212735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225020	49.879	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.760%
35) Dibromofluoromethane	8.024	113	174419	50.083	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.160%
50) Toluene-d8	10.443	98	750364	53.849	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.700%
62) 4-Bromofluorobenzene	12.733	95	269827	51.324	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.640%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	4287	2.546	ug/l	99
44) Trichloroethene	9.217	130	9065	2.326	ug/l	93
64) Tetrachloroethene	10.979	164	397110	105.997	ug/l	98
85) sec-Butylbenzene	13.503	105	10278	0.578	ug/l	99
90) Hexachloroethane	14.211	117	3862	1.289	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069222.D
Acq On : 27 Oct 2021 16:47
Operator : JC/MD
Sample : M4338-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5D

Quant Time: Oct 28 05:21:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

