

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069223.D
 Acq On : 27 Oct 2021 17:12
 Operator : JC/MD
 Sample : M4338-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3D(ONSITE)

Quant Time: Oct 28 05:22:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	360405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	630389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	599323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	218314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	237138	49.905	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 99.800%			
35) Dibromofluoromethane	8.024	113	185389	50.547	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.100%			
50) Toluene-d8	10.443	98	783534	53.393	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.780%			
62) 4-Bromofluorobenzene	12.734	95	275747	49.804	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.600%			
Target Compounds						
44) Trichloroethene	9.220	130	2919	0.711	ug/l	88
64) Tetrachloroethene	10.980	164	7056	1.828	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069223.D
Acq On : 27 Oct 2021 17:12
Operator : JC/MD
Sample : M4338-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3D(ONSITE)

Quant Time: Oct 28 05:22:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

