

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069226.D
 Acq On : 27 Oct 2021 18:28
 Operator : JC/MD
 Sample : M4338-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2S(ONSITE)

Quant Time: Oct 28 05:22:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	347240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	602154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	587226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	214412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	230490	50.345	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.680%		
35) Dibromofluoromethane	8.024	113	177584	50.689	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 101.380%		
50) Toluene-d8	10.443	98	756410	53.961	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 107.920%		
62) 4-Bromofluorobenzene	12.733	95	269422	50.943	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 101.880%		
Target Compounds						
64) Tetrachloroethene	10.979	164	32944	8.709	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069226.D
Acq On : 27 Oct 2021 18:28
Operator : JC/MD
Sample : M4338-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2S(ONSITE)

Quant Time: Oct 28 05:22:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

