

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069240.D
 Acq On : 28 Oct 2021 1:10
 Operator : JC/MD
 Sample : M4338-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-15S

Quant Time: Oct 28 11:01:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	329274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	578075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	558393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	204029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	220194	50.720	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.440%
35) Dibromofluoromethane	8.024	113	167610	49.835	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.680%
50) Toluene-d8	10.443	98	721765	53.635	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.260%
62) 4-Bromofluorobenzene	12.734	95	256389	50.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.000%
Target Compounds						
64) Tetrachloroethene	10.980	164	39977	11.114	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069240.D
Acq On : 28 Oct 2021 1:10
Operator : JC/MD
Sample : M4338-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-15S

Quant Time: Oct 28 11:01:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

