

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069212.D
 Acq On : 27 Oct 2021 12:25
 Operator : JC/MD
 Sample : VN1027WBL03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1027WBL03

Quant Time: Oct 28 05:20:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	334684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	585348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	560566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	206672	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222504	50.424	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.840%
35) Dibromofluoromethane	8.024	113	172608	50.683	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.360%
50) Toluene-d8	10.443	98	737138	54.096	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.200%
62) 4-Bromofluorobenzene	12.733	95	257900	50.165	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.320%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069212.D
Acq On : 27 Oct 2021 12:25
Operator : JC/MD
Sample : VN1027WBL03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1027WBL03

Quant Time: Oct 28 05:20:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.002 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25




Instrument :
MSVOA_N
ClientSampleId :
VN1027WBL03


Tgt Ion:168 Resp: 334684
Ion Ratio Lower Upper
168 100
99 62.4 48.9 73.3



#33
1,2-Dichloroethane-d4
Concen: 50.424 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. 0.000 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25

Tgt Ion: 65 Resp: 222504
Ion Ratio Lower Upper
65 100
67 51.6 0.0 103.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.971 min Scan# 2603
Delta R.T. 0.003 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25




Instrument :
MSVOA_N
ClientSampleId :
VN1027WBL03


Tgt Ion:114 Resp: 585348
Ion Ratio Lower Upper
114 100
63 20.8 0.0 42.6
88 15.9 0.0 32.0



#35
Dibromofluoromethane
Concen: 50.683 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.003 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25

Tgt Ion:113 Resp: 172608
Ion Ratio Lower Upper
113 100
111 102.4 83.2 124.8
192 17.9 15.0 22.4









#50
Toluene-d8
Concen: 54.096 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25




Instrument :
MSVOA_N
ClientSampleId :
VN1027WBL03


Tgt Ion: 98 Resp: 737138
Ion Ratio Lower Upper
98 100
100 63.8 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 50.165 ug/l
RT: 12.733 min Scan# 4006
Delta R.T. 0.000 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25

Tgt Ion: 95 Resp: 257900
Ion Ratio Lower Upper
95 100
174 75.0 0.0 144.8
176 71.1 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.746 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25

Instrument :
MSVOA_N
ClientSampleId :
VN1027WBL03

Tgt Ion:117 Resp: 560566
Ion Ratio Lower Upper
117 100
82 56.6 46.4 69.6
119 32.5 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.003 min
Lab File: VN069212.D
Acq: 27 Oct 2021 12:25

Tgt Ion:152 Resp: 206672
Ion Ratio Lower Upper
152 100
115 63.2 31.4 94.3
150 155.9 0.0 351.0

