

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102721\  
 Data File : VN069237.D  
 Acq On : 27 Oct 2021 23:55  
 Operator : JC/MD  
 Sample : VN1027WBS06  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 31 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1027WBS06

Manual Integrations  
 APPROVED

MMDadoda  
 10/29/2021 1:25:47 PM

Quant Time: Oct 28 01:25:18 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 13 05:44:40 2021  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.088  | 168  | 216248   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.971  | 114  | 397116   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.747 | 117  | 360393   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.678 | 152  | 141232   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |            |      |
|-----------------------------|--------|-------|----------|----------|------------|------|
| System Monitoring Compounds |        |       |          |          |            |      |
| 33) 1,2-Dichloroethane-d4   | 8.437  | 65    | 171774   | 60.247   | ug/l       | 0.00 |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 120.500% |      |
| 35) Dibromofluoromethane    | 8.024  | 113   | 129713   | 56.142   | ug/l       | 0.00 |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 112.280% |      |
| 50) Toluene-d8              | 10.443 | 98    | 503811   | 54.498   | ug/l       | 0.00 |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 109.000% |      |
| 62) 4-Bromofluorobenzene    | 12.734 | 95    | 174513   | 50.035   | ug/l       | 0.00 |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 100.060% |      |

|                              |       |     |        |         |        |     |
|------------------------------|-------|-----|--------|---------|--------|-----|
| Target Compounds             |       |     |        |         | Qvalue |     |
| 2) Dichlorodifluoromethane   | 2.073 | 85  | 43603  | 17.033  | ug/l   | 99  |
| 3) Chloromethane             | 2.303 | 50  | 48038  | 18.175  | ug/l   | 99  |
| 4) Vinyl Chloride            | 2.443 | 62  | 57461  | 21.403  | ug/l   | 97  |
| 5) Bromomethane              | 2.848 | 94  | 39731  | 26.905  | ug/l   | 100 |
| 6) Chloroethane              | 3.014 | 64  | 38862  | 24.563  | ug/l   | 98  |
| 7) Trichlorofluoromethane    | 3.371 | 101 | 78057  | 20.764  | ug/l   | 98  |
| 8) Diethyl Ether             | 3.829 | 74  | 28792  | 20.533  | ug/l   | 99  |
| 9) 1,1,2-Trichlorotrifluo... | 4.213 | 101 | 41302  | 20.433  | ug/l   | 98  |
| 10) Methyl Iodide            | 4.427 | 142 | 49081  | 16.729  | ug/l   | 97  |
| 11) Tert butyl alcohol       | 5.366 | 59  | 63283  | 121.469 | ug/l # | 86  |
| 12) 1,1-Dichloroethene       | 4.186 | 96  | 39662  | 18.981  | ug/l   | 98  |
| 13) Acrolein                 | 4.049 | 56  | 15589  | 67.641  | ug/l   | 97  |
| 14) Allyl chloride           | 4.846 | 41  | 69136  | 18.457  | ug/l   | 98  |
| 15) Acrylonitrile            | 5.557 | 53  | 148621 | 127.312 | ug/l   | 99  |
| 16) Acetone                  | 4.288 | 43  | 128994 | 121.235 | ug/l   | 100 |
| 17) Carbon Disulfide         | 4.537 | 76  | 79736  | 14.061  | ug/l   | 99  |
| 18) Methyl Acetate           | 4.857 | 43  | 104189 | 25.817  | ug/l   | 98  |
| 19) Methyl tert-butyl Ether  | 5.618 | 73  | 164741 | 21.174  | ug/l   | 99  |
| 20) Methylene Chloride       | 5.098 | 84  | 50832  | 20.757  | ug/l   | 92  |
| 21) trans-1,2-Dichloroethene | 5.602 | 96  | 42000  | 18.605  | ug/l   | 95  |
| 22) Diisopropyl ether        | 6.498 | 45  | 155702 | 19.988  | ug/l   | 97  |
| 23) Vinyl Acetate            | 6.436 | 43  | 596569 | 101.305 | ug/l   | 99  |
| 24) 1,1-Dichloroethane       | 6.388 | 63  | 87805  | 20.747  | ug/l   | 99  |
| 25) 2-Butanone               | 7.337 | 43  | 205911 | 127.979 | ug/l   | 98  |
| 26) 2,2-Dichloropropane      | 7.329 | 77  | 60746  | 15.061  | ug/l   | 98  |
| 27) cis-1,2-Dichloroethene   | 7.332 | 96  | 54085  | 19.820  | ug/l   | 97  |
| 28) Bromochloromethane       | 7.662 | 49  | 32828  | 17.126  | ug/l   | 97  |
| 29) Tetrahydrofuran          | 7.686 | 42  | 134215 | 128.604 | ug/l   | 97  |
| 30) Chloroform               | 7.823 | 83  | 93928  | 21.059  | ug/l   | 96  |
| 31) Cyclohexane              | 8.099 | 56  | 76671  | 17.295  | ug/l   | 95  |
| 32) 1,1,1-Trichloroethane    | 8.019 | 97  | 83215  | 20.013  | ug/l   | 99  |
| 36) 1,1-Dichloropropene      | 8.225 | 75  | 62612  | 18.459  | ug/l   | 99  |
| 37) Ethyl Acetate            | 7.412 | 43  | 80289  | 22.919  | ug/l   | 99  |
| 38) Carbon Tetrachloride     | 8.212 | 117 | 67145  | 18.024  | ug/l   | 98  |
| 39) Methylcyclohexane        | 9.462 | 83  | 72902  | 16.698  | ug/l   | 98  |
| 40) Benzene                  | 8.461 | 78  | 203927 | 19.743  | ug/l   | 99  |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102721\  
 Data File : VN069237.D  
 Acq On : 27 Oct 2021 23:55  
 Operator : JC/MD  
 Sample : VN1027WBS06  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 31 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1027WBS06

Manual Integrations  
 APPROVED

MMDadoda  
 10/29/2021 1:25:47 PM

Quant Time: Oct 28 01:25:18 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 13 05:44:40 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.640  | 41   | 39812    | 22.545  | ug/l   | 94       |
| 42) 1,2-Dichloroethane        | 8.531  | 62   | 78424    | 21.441  | ug/l   | 100      |
| 43) Isopropyl Acetate         | 8.563  | 43   | 131339   | 21.248  | ug/l   | 99       |
| 44) Trichloroethene           | 9.217  | 130  | 49296    | 19.070  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 9.494  | 63   | 52965    | 19.762  | ug/l   | 95       |
| 46) Dibromomethane            | 9.580  | 93   | 37146    | 20.304  | ug/l   | 99       |
| 47) Bromodichloromethane      | 9.765  | 83   | 70393    | 18.567  | ug/l   | 99       |
| 48) Methyl methacrylate       | 9.561  | 41   | 61342    | 21.121  | ug/l   | 97       |
| 49) 1,4-Dioxane               | 9.571  | 88   | 26102    | 480.328 | ug/l   | 93       |
| 51) 4-Methyl-2-Pentanone      | 10.333 | 43   | 422800   | 117.967 | ug/l   | 99       |
| 52) Toluene                   | 10.507 | 92   | 130133   | 19.583  | ug/l   | 97       |
| 53) t-1,3-Dichloropropene     | 10.722 | 75   | 71458    | 16.639  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 10.191 | 75   | 77563    | 17.359  | ug/l   | 99       |
| 55) 1,1,2-Trichloroethane     | 10.902 | 97   | 54292    | 20.722  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 10.762 | 69   | 86746    | 19.720  | ug/l   | 99       |
| 57) 1,3-Dichloropropane       | 11.047 | 76   | 91726    | 21.232  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 10.044 | 63   | 114233   | 100.412 | ug/l   | 98       |
| 59) 2-Hexanone                | 11.087 | 43   | 310050   | 120.631 | ug/l   | 99       |
| 60) Dibromochloromethane      | 11.240 | 129  | 51895    | 17.548  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 11.347 | 107  | 55487    | 20.647  | ug/l   | 100      |
| 64) Tetrachloroethene         | 10.980 | 164  | 44484    | 19.161  | ug/l   | 99       |
| 65) Chlorobenzene             | 11.773 | 112  | 137171   | 20.114  | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane | 11.843 | 131  | 51404    | 19.258  | ug/l   | 99       |
| 67) Ethyl Benzene             | 11.846 | 91   | 248342   | 19.604  | ug/l   | 99       |
| 68) m/p-Xylenes               | 11.956 | 106  | 184143   | 38.346  | ug/l   | 98       |
| 69) o-Xylene                  | 12.283 | 106  | 94780    | 19.718  | ug/l   | 100      |
| 70) Styrene                   | 12.296 | 104  | 151165   | 19.319  | ug/l   | 100      |
| 71) Bromoform                 | 12.460 | 173  | 35400    | 16.556  | ug/l # | 99       |
| 73) Isopropylbenzene          | 12.581 | 105  | 244531   | 20.801  | ug/l   | 99       |
| 74) N-amyl acetate            | 12.390 | 43   | 92581    | 20.788  | ug/l   | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 12.830 | 83   | 84407    | 24.846  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 12.881 | 75   | 81058m   | 26.974  | ug/l   |          |
| 77) Bromobenzene              | 12.862 | 156  | 56574    | 21.764  | ug/l   | 99       |
| 78) n-propylbenzene           | 12.921 | 91   | 263277   | 19.929  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 13.010 | 91   | 170572   | 20.948  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene    | 13.061 | 105  | 198354   | 20.258  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 12.629 | 75   | 19428    | 15.468  | ug/l   | 86       |
| 82) 4-Chlorotoluene           | 13.109 | 91   | 157476   | 20.014  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 13.326 | 119  | 180334   | 21.157  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 13.372 | 105  | 192467   | 20.401  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 13.503 | 105  | 236953   | 20.073  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.619 | 119  | 182928   | 19.320  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 13.619 | 146  | 90953    | 20.228  | ug/l   | 100      |
| 88) 1,4-Dichlorobenzene       | 13.696 | 146  | 88866    | 19.913  | ug/l   | 98       |
| 89) n-Butylbenzene            | 13.946 | 91   | 135146   | 17.145  | ug/l   | 99       |
| 90) Hexachloroethane          | 14.211 | 117  | 32854    | 16.518  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 13.991 | 146  | 91485    | 20.945  | ug/l   | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 14.603 | 75   | 15355    | 23.443  | ug/l   | 94       |
| 93) 1,2,4-Trichlorobenzene    | 15.268 | 180  | 37159    | 17.452  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 15.373 | 225  | 22075    | 17.749  | ug/l   | 97       |
| 95) Naphthalene               | 15.509 | 128  | 110628   | 19.404  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.702 | 180  | 38258    | 18.613  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102721\  
Data File : VN069237.D  
Acq On : 27 Oct 2021 23:55  
Operator : JC/MD  
Sample : VN1027WBS06  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 31 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM

Quant Time: Oct 28 01:25:18 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 13 05:44:40 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.088 min Scan# 2274  
Delta R.T. 0.002 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 168 Resp: 216248  
Ion Ratio Lower Upper  
168 100  
99 61.3 48.9 73.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#2  
Dichlorodifluoromethane  
Concen: 17.033 ug/l  
RT: 2.073 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 85 Resp: 43603  
Ion Ratio Lower Upper  
85 100  
87 32.9 16.3 48.8








#3  
Chloromethane  
Concen: 18.175 ug/l  
RT: 2.303 min Scan# 117  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 50 Resp: 48038  
Ion Ratio Lower Upper  
50 100  
52 33.2 26.3 39.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#4  
Vinyl Chloride  
Concen: 21.403 ug/l  
RT: 2.443 min Scan# 169  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


Tgt Ion: 62 Resp: 57461  
Ion Ratio Lower Upper  
62 100  
64 33.6 25.7 38.5





#5  
Bromomethane  
Concen: 26.905 ug/l  
RT: 2.848 min Scan# 320  
Delta R.T. 0.011 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06



Tgt Ion: 94 Resp: 39731  
Ion Ratio Lower Upper  
94 100  
96 92.9 74.2 111.4




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#6  
Chloroethane  
Concen: 24.563 ug/l  
RT: 3.014 min Scan# 382  
Delta R.T. 0.005 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 64 Resp: 38862  
Ion Ratio Lower Upper  
64 100  
66 33.0 25.6 38.4









#7  
Trichlorofluoromethane  
Concen: 20.764 ug/l  
RT: 3.371 min Scan# 515  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion:101 Resp: 78057  
Ion Ratio Lower Upper  
101 100  
103 62.9 51.7 77.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM




#8  
Diethyl Ether  
Concen: 20.533 ug/l  
RT: 3.829 min Scan# 686  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 74 Resp: 28792  
Ion Ratio Lower Upper  
74 100  
45 93.1 47.0 141.2









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.433 ug/l  
RT: 4.213 min Scan# 829  
Delta R.T. -0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55





Tgt Ion:101 Resp: 41302  
Ion Ratio Lower Upper  
101 100  
85 45.1 37.2 55.8  
151 74.6 61.1 91.7



#10  
Methyl Iodide  
Concen: 16.729 ug/l  
RT: 4.427 min Scan# 909  
Delta R.T. 0.005 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55



Tgt Ion:142 Resp: 49081  
Ion Ratio Lower Upper  
142 100  
127 45.1 34.4 51.6  
141 14.0 11.3 16.9



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM






#11  
Tert butyl alcohol  
Concen: 121.469 ug/l  
RT: 5.366 min Scan# 1259  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 59 Resp: 63283  
Ion Ratio Lower Upper  
59 100  
57 5.2 8.3 12.5#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#12  
1,1-Dichloroethene  
Concen: 18.981 ug/l  
RT: 4.186 min Scan# 819  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 96 Resp: 39662  
Ion Ratio Lower Upper  
96 100  
61 158.7 129.8 194.6  
98 63.6 51.4 77.2









#13  
Acrolein  
Concen: 67.641 ug/l  
RT: 4.049 min Scan# 768  
Delta R.T. 0.008 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 56 Resp: 15589  
Ion Ratio Lower Upper  
56 100  
55 69.9 58.1 87.1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#14  
Allyl chloride  
Concen: 18.457 ug/l  
RT: 4.846 min Scan# 1065  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 41 Resp: 69136  
Ion Ratio Lower Upper  
41 100  
39 71.8 55.6 83.4  
76 33.8 26.4 39.6









#15  
Acrylonitrile  
Concen: 127.312 ug/l  
RT: 5.557 min Scan# 1330  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 53 Resp: 148621  
Ion Ratio Lower Upper  
53 100  
52 82.9 65.7 98.5  
51 36.3 28.9 43.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#16  
Acetone  
Concen: 121.235 ug/l  
RT: 4.288 min Scan# 857  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 43 Resp: 128994  
Ion Ratio Lower Upper  
43 100  
58 30.8 24.5 36.7









#17  
Carbon Disulfide  
Concen: 14.061 ug/l  
RT: 4.537 min Scan# 950  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 76 Resp: 79736  
Ion Ratio Lower Upper  
76 100  
78 9.4 7.2 10.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#18  
Methyl Acetate  
Concen: 25.817 ug/l  
RT: 4.857 min Scan# 1069  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 43 Resp: 104189  
Ion Ratio Lower Upper  
43 100  
74 24.6 18.9 28.3





#19





Methyl tert-butyl Ether  
 Concen: 21.174 ug/l  
 RT: 5.618 min Scan# 1353  
 Delta R.T. 0.003 min  
 Lab File: VN069237.D  
 Acq: 27 Oct 2021 23:55

Tgt Ion: 73 Resp: 164741  
 Ion Ratio Lower Upper  
 73 100  
 57 22.2 17.3 25.9

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1027WBS06

Manual Integrations  
 APPROVED




MMDadoda  
 10/29/2021 1:25:47 PM



#20

Methylene Chloride  
 Concen: 20.757 ug/l  
 RT: 5.098 min Scan# 1159  
 Delta R.T. 0.003 min  
 Lab File: VN069237.D  
 Acq: 27 Oct 2021 23:55

Tgt Ion: 84 Resp: 50832  
 Ion Ratio Lower Upper  
 84 100  
 49 111.6 97.6 146.4  
 51 34.3 30.3 45.5  
 86 60.1 51.8 77.6





#21





trans-1,2-Dichloroethene  
Concen: 18.605 ug/l  
RT: 5.602 min Scan# 1347  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 96 Resp: 42000  
Ion Ratio Lower Upper  
96 100  
61 132.4 112.4 168.6  
98 63.0 52.0 78.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED




MMDadoda  
10/29/2021 1:25:47 PM




#22

Diisopropyl ether  
Concen: 19.988 ug/l  
RT: 6.498 min Scan# 1681  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 45 Resp: 155702  
Ion Ratio Lower Upper  
45 100  
43 62.0 48.0 72.0  
87 30.8 23.4 35.0  
59 13.4 9.5 14.3






#23  
 Vinyl Acetate  
 Concen: 101.305 ug/l  
 RT: 6.436 min Scan# 1658  
 Delta R.T. 0.003 min  
 Lab File: VN069237.D  
 Acq: 27 Oct 2021 23:55




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1027WBS06

Manual Integrations  
 APPROVED

MMDadoda  
 10/29/2021 1:25:47 PM






Tgt Ion: 43 Resp: 596569  
 Ion Ratio Lower Upper  
 43 100  
 86 10.9 8.3 12.5



#24  
 1,1-Dichloroethane  
 Concen: 20.747 ug/l  
 RT: 6.388 min Scan# 1640  
 Delta R.T. 0.000 min  
 Lab File: VN069237.D  
 Acq: 27 Oct 2021 23:55

Tgt Ion: 63 Resp: 87805  
 Ion Ratio Lower Upper  
 63 100  
 98 6.6 3.5 10.4  
 100 4.5 2.1 6.5









#25  
2-Butanone  
Concen: 127.979 ug/l  
RT: 7.337 min Scan# 1994  
Delta R.T. 0.005 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 43 Resp: 205911  
Ion Ratio Lower Upper  
43 100  
72 27.7 21.2 31.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#26  
2,2-Dichloropropane  
Concen: 15.061 ug/l  
RT: 7.329 min Scan# 1991  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 77 Resp: 60746  
Ion Ratio Lower Upper  
77 100  
97 23.5 11.2 33.6








#27  
cis-1,2-Dichloroethene  
Concen: 19.820 ug/l  
RT: 7.332 min Scan# 1992  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 96 Resp: 54085  
Ion Ratio Lower Upper  
96 100  
61 140.7 0.0 290.6  
98 63.5 0.0 127.2

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#28  
Bromochloromethane  
Concen: 17.126 ug/l  
RT: 7.662 min Scan# 2115  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 49 Resp: 32828  
Ion Ratio Lower Upper  
49 100  
129 1.1 0.0 4.4  
130 78.9 60.8 91.2









#29  
Tetrahydrofuran  
Concen: 128.604 ug/l  
RT: 7.686 min Scan# 2124  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 42 Resp: 134215  
Ion Ratio Lower Upper  
42 100  
72 46.4 35.4 53.2  
71 44.3 33.8 50.6

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#30  
Chloroform  
Concen: 21.059 ug/l  
RT: 7.823 min Scan# 2175  
Delta R.T. 0.005 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 83 Resp: 93928  
Ion Ratio Lower Upper  
83 100  
85 67.0 51.1 76.7









#31  
Cyclohexane  
Concen: 17.295 ug/l  
RT: 8.099 min Scan# 2278  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 56    | Resp: | 76671 |
| Ion Ratio | Lower | Upper |       |
| 56        | 100   |       |       |
| 69        | 36.9  | 26.9  | 40.3  |
| 84        | 89.0  | 67.6  | 101.4 |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#32  
1,1,1-Trichloroethane  
Concen: 20.013 ug/l  
RT: 8.019 min Scan# 2248  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 97    | Resp: | 83215 |
| Ion Ratio | Lower | Upper |       |
| 97        | 100   |       |       |
| 99        | 62.6  | 50.8  | 76.2  |
| 61        | 44.1  | 34.6  | 51.8  |









#33  
1,2-Dichloroethane-d4  
Concen: 60.247 ug/l  
RT: 8.437 min Scan# 2404  
Delta R.T. -0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Tgt Ion: 65 Resp: 171774  
Ion Ratio Lower Upper  
65 100  
67 52.3 0.0 103.8




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.971 min Scan# 2603  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 114 Resp: 397116  
Ion Ratio Lower Upper  
114 100  
63 20.3 0.0 42.6  
88 15.7 0.0 32.0









#35  
Dibromofluoromethane  
Concen: 56.142 ug/l  
RT: 8.024 min Scan# 2250  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55





Tgt Ion: 113 Resp: 129713  
Ion Ratio Lower Upper  
113 100  
111 102.3 83.2 124.8  
192 17.9 15.0 22.4



#36  
1,1-Dichloropropene  
Concen: 18.459 ug/l  
RT: 8.225 min Scan# 2325  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion: 75 Resp: 62612  
Ion Ratio Lower Upper  
75 100  
110 36.0 17.9 53.8  
77 31.9 24.7 37.1



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM







#37  
Ethyl Acetate  
Concen: 22.919 ug/l  
RT: 7.412 min Scan# 2022  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 43    | Resp: | 80289 |
| Ion      | Ratio | Lower | Upper |
| 43       | 100   |       |       |
| 61       | 14.7  | 11.2  | 16.8  |
| 70       | 10.4  | 8.4   | 12.6  |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#38  
Carbon Tetrachloride  
Concen: 18.024 ug/l  
RT: 8.212 min Scan# 2320  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 117   | Resp: | 67145 |
| Ion      | Ratio | Lower | Upper |
| 117      | 100   |       |       |
| 119      | 98.5  | 76.9  | 115.3 |
| 121      | 29.6  | 24.6  | 37.0  |





#39

Methylcyclohexane

Concen: 16.698 ug/l

RT: 9.462 min Scan# 2786

Delta R.T. 0.000 min

Lab File: VN069237.D

Acq: 27 Oct 2021 23:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 72902 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 55       | 73.4  | 61.4  | 92.0  |
| 98       | 45.6  | 36.4  | 54.6  |

Instrument :





MSVOA\_N

ClientSampleId :

VN1027WBS06

Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#40

Benzene

Concen: 19.743 ug/l




RT: 8.461 min Scan# 2413


Delta R.T. 0.000 min

Lab File: VN069237.D

Acq: 27 Oct 2021 23:55

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 78    | Resp: | 203927 |
| Ion      | Ratio | Lower | Upper  |
| 78       | 100   |       |        |
| 77       | 23.6  | 19.1  | 28.7   |





#41  
Methacrylonitrile  
Concen: 22.545 ug/l  
RT: 7.640 min Scan# 2107  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55





Tgt Ion: 41 Resp: 39812

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 39  | 60.5  | 44.9  | 67.3  |
| 67  | 82.9  | 62.6  | 93.8  |
| 52  | 35.6  | 26.2  | 39.4  |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED




MMDadoda  
10/29/2021 1:25:47 PM




#42  
1,2-Dichloroethane  
Concen: 21.441 ug/l  
RT: 8.531 min Scan# 2439  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 62 Resp: 78424

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 62  | 100   |       |       |
| 98  | 8.4   | 0.0   | 17.2  |









#43  
Isopropyl Acetate  
Concen: 21.248 ug/l  
RT: 8.563 min Scan# 2451  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 43 Resp: 131339  
Ion Ratio Lower Upper  
43 100  
61 26.2 20.4 30.6  
87 14.0 10.7 16.1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#44  
Trichloroethene  
Concen: 19.070 ug/l  
RT: 9.217 min Scan# 2695  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion:130 Resp: 49296  
Ion Ratio Lower Upper  
130 100  
95 103.7 0.0 208.0





#45





1,2-Dichloropropane  
Concen: 19.762 ug/l  
RT: 9.494 min Scan# 2798  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 63 Resp: 52965  
Ion Ratio Lower Upper  
63 100  
65 29.0 25.6 38.4

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED




MMDadoda  
10/29/2021 1:25:47 PM



#46

Dibromomethane  
Concen: 20.304 ug/l  
RT: 9.580 min Scan# 2830  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 93 Resp: 37146  
Ion Ratio Lower Upper  
93 100  
95 82.8 66.5 99.7  
174 91.5 73.8 110.8









#47  
Bromodichloromethane  
Concen: 18.567 ug/l  
RT: 9.765 min Scan# 2899  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 83 Resp: 70393  
Ion Ratio Lower Upper  
83 100  
85 63.9 50.7 76.1  
127 8.8 6.5 9.7

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#48  
Methyl methacrylate  
Concen: 21.121 ug/l  
RT: 9.561 min Scan# 2823  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 41 Resp: 61342  
Ion Ratio Lower Upper  
41 100  
69 84.7 66.4 99.6  
39 65.7 50.2 75.2









#49  
1,4-Dioxane  
Concen: 480.328 ug/l  
RT: 9.571 min Scan# 2827  
Delta R.T. 0.002 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 88 Resp: 26102  
Ion Ratio Lower Upper  
88 100  
43 29.8 24.9 37.3  
58 63.8 56.9 85.3

Instrument :  
MSVOA\_N  
ClientSampled :  
VN1027WBS06




Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#50  
Toluene-d8  
Concen: 54.498 ug/l  
RT: 10.443 min Scan# 3152  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 98 Resp: 503811  
Ion Ratio Lower Upper  
98 100  
100 64.1 51.5 77.3









#51  
4-Methyl-2-Pentanone  
Concen: 117.967 ug/l  
RT: 10.333 min Scan# 3111  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 43 Resp: 422800  
Ion Ratio Lower Upper  
43 100  
58 39.1 31.0 46.4

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#52  
Toluene  
Concen: 19.583 ug/l  
RT: 10.507 min Scan# 3176  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 92 Resp: 130133  
Ion Ratio Lower Upper  
92 100  
91 176.7 138.5 207.7









#53  
t-1,3-Dichloropropene  
Concen: 16.639 ug/l  
RT: 10.722 min Scan# 3256  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 75 Resp: 71458  
Ion Ratio Lower Upper  
75 100  
77 31.0 25.4 38.2

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#54  
cis-1,3-Dichloropropene  
Concen: 17.359 ug/l  
RT: 10.191 min Scan# 3058  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 75 Resp: 77563  
Ion Ratio Lower Upper  
75 100  
77 32.4 25.6 38.4  
39 53.9 42.9 64.3









#55  
1,1,2-Trichloroethane  
Concen: 20.722 ug/l  
RT: 10.902 min Scan# 3323  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 97    | Resp: | 54292 |
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 83       | 84.2  | 70.5  | 105.7 |
| 85       | 54.4  | 45.3  | 67.9  |
| 99       | 61.6  | 50.7  | 76.1  |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#56  
Ethyl methacrylate  
Concen: 19.720 ug/l  
RT: 10.762 min Scan# 3271  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 69    | Resp: | 86746 |
| Ion      | Ratio | Lower | Upper |
| 69       | 100   |       |       |
| 41       | 66.9  | 53.3  | 79.9  |
| 39       | 43.0  | 33.3  | 49.9  |





#57  
1,3-Dichloropropane  
Concen: 21.232 ug/l  
RT: 11.047 min Scan# 3377  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion: 76 Resp: 91726  
Ion Ratio Lower Upper  
76 100  
78 32.3 25.8 38.8

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#58  
2-Chloroethyl Vinyl ether  
Concen: 100.412 ug/l  
RT: 10.044 min Scan# 3003  
Delta R.T. 0.001 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55



Tgt Ion: 63 Resp: 114233  
Ion Ratio Lower Upper  
63 100  
106 28.5 22.2 33.2









#59  
2-Hexanone  
Concen: 120.631 ug/l  
RT: 11.087 min Scan# 3392  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 43 Resp: 310050  
Ion Ratio Lower Upper  
43 100  
58 54.4 26.9 80.7

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#60  
Dibromochloromethane  
Concen: 17.548 ug/l  
RT: 11.240 min Scan# 3449  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


Tgt Ion:129 Resp: 51895  
Ion Ratio Lower Upper  
129 100  
127 76.9 39.1 117.2





#61  
1,2-Dibromoethane  
Concen: 20.647 ug/l  
RT: 11.347 min Scan# 3489  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:107 Resp: 55487  
Ion Ratio Lower Upper  
107 100  
109 92.9 74.7 112.1

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#62  
4-Bromofluorobenzene  
Concen: 50.035 ug/l  
RT: 12.734 min Scan# 4006  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion: 95 Resp: 174513  
Ion Ratio Lower Upper  
95 100  
174 75.3 0.0 144.8  
176 72.4 0.0 140.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.747 min Scan# 3638  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:117 Resp: 360393  
Ion Ratio Lower Upper  
117 100  
82 57.0 46.4 69.6  
119 32.0 26.0 39.0

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#64  
Tetrachloroethene  
Concen: 19.161 ug/l  
RT: 10.980 min Scan# 3352  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion:164 Resp: 44484  
Ion Ratio Lower Upper  
164 100  
166 124.8 99.8 149.8  
129 97.4 79.2 118.8  
131 94.8 76.3 114.5





#65  
Chlorobenzene  
Concen: 20.114 ug/l  
RT: 11.773 min Scan# 3648  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:112 Resp: 137171  
Ion Ratio Lower Upper  
112 100  
114 30.4 25.5 38.3

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.258 ug/l  
RT: 11.843 min Scan# 3674  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55



Tgt Ion:131 Resp: 51404  
Ion Ratio Lower Upper  
131 100  
133 96.0 47.3 142.0  
119 65.2 33.0 99.0









#67  
Ethyl Benzene  
Concen: 19.604 ug/l  
RT: 11.846 min Scan# 3675  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 91 Resp: 248342  
Ion Ratio Lower Upper  
91 100  
106 31.3 24.8 37.2

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#68  
m/p-Xylenes  
Concen: 38.346 ug/l  
RT: 11.956 min Scan# 3716  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 106 Resp: 184143  
Ion Ratio Lower Upper  
106 100  
91 211.0 166.8 250.2









#69  
o-Xylene  
Concen: 19.718 ug/l  
RT: 12.283 min Scan# 3838  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion:106 Resp: 94780  
Ion Ratio Lower Upper  
106 100  
91 219.0 109.6 328.6

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#70  
Styrene  
Concen: 19.319 ug/l  
RT: 12.296 min Scan# 3843  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion:104 Resp: 151165  
Ion Ratio Lower Upper  
104 100  
78 52.4 42.2 63.4  
103 55.2 44.4 66.6









#71  
Bromoform  
Concen: 16.556 ug/l  
RT: 12.460 min Scan# 3904  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion:173 Resp: 35400  
Ion Ratio Lower Upper  
173 100  
175 48.4 23.8 71.5  
254 0.0 0.0 0.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.678 min Scan# 4358  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


Tgt Ion:152 Resp: 141232  
Ion Ratio Lower Upper  
152 100  
115 63.8 31.4 94.3  
150 164.6 0.0 351.0





#73  
Isopropylbenzene  
Concen: 20.801 ug/l  
RT: 12.581 min Scan# 3949  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:105 Resp: 244531  
Ion Ratio Lower Upper  
105 100  
120 25.7 13.1 39.3

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#74  
N-ethyl acetate  
Concen: 20.788 ug/l  
RT: 12.390 min Scan# 3878  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55



Tgt Ion: 43 Resp: 92581  
Ion Ratio Lower Upper  
43 100  
70 42.3 32.2 48.2  
55 23.6 18.0 27.0  
61 24.1 19.0 28.4









#75  
1,1,2,2-Tetrachloroethane  
Concen: 24.846 ug/l  
RT: 12.830 min Scan# 4042  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 83 Resp: 84407  
Ion Ratio Lower Upper  
83 100  
131 10.0 5.1 15.2  
85 64.5 32.3 96.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#76  
1,2,3-Trichloropropane  
Concen: 26.974 ug/l m  
RT: 12.881 min Scan# 4061  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


Tgt Ion: 75 Resp: 81058  
Ion Ratio Lower Upper  
75 100  
77 154.7 96.2 288.4





#77  
Bromobenzene  
Concen: 21.764 ug/l  
RT: 12.862 min Scan# 4054  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:156 Resp: 56574  
Ion Ratio Lower Upper  
156 100  
77 221.7 111.8 335.3  
158 95.1 48.4 145.0

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM



#78  
n-propylbenzene  
Concen: 19.929 ug/l  
RT: 12.921 min Scan# 4076  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55



Tgt Ion: 91 Resp: 263277  
Ion Ratio Lower Upper  
91 100  
120 22.3 11.4 34.1









#79  
2-Chlorotoluene  
Concen: 20.948 ug/l  
RT: 13.010 min Scan# 4109  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion: 91 Resp: 170572  
Ion Ratio Lower Upper  
91 100  
126 32.8 16.4 49.4

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 20.258 ug/l  
RT: 13.061 min Scan# 4128  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


Tgt Ion: 105 Resp: 198354  
Ion Ratio Lower Upper  
105 100  
120 46.8 23.4 70.2





#81  
trans-1,4-Dichloro-2-butene  
Concen: 15.468 ug/l  
RT: 12.629 min Scan# 3967  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion: 75 Resp: 19428  
Ion Ratio Lower Upper  
75 100  
53 112.5 75.4 113.2  
89 47.9 36.9 55.3

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM



#82  
4-Chlorotoluene  
Concen: 20.014 ug/l  
RT: 13.109 min Scan# 4146  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion: 91 Resp: 157476  
Ion Ratio Lower Upper  
91 100  
126 32.4 16.3 48.8





#83  
tert-Butylbenzene  
Concen: 21.157 ug/l  
RT: 13.326 min Scan# 4227  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:119 Resp: 180334  
Ion Ratio Lower Upper  
119 100  
91 68.0 34.6 103.9  
134 25.6 12.6 37.6

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#84  
1,2,4-Trimethylbenzene  
Concen: 20.401 ug/l  
RT: 13.372 min Scan# 4244  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion:105 Resp: 192467  
Ion Ratio Lower Upper  
105 100  
120 43.9 21.9 65.5





#85  
 sec-Butylbenzene  
 Concen: 20.073 ug/l  
 RT: 13.503 min Scan# 4293  
 Delta R.T. 0.000 min  
 Lab File: VN069237.D  
 Acq: 27 Oct 2021 23:55




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1027WBS06




Tgt Ion:105 Resp: 236953  
 Ion Ratio Lower Upper  
 105 100  
 134 18.9 9.8 29.4

Manual Integrations  
 APPROVED



MMDadoda  
 10/29/2021 1:25:47 PM




#86  
 p-Isopropyltoluene  
 Concen: 19.320 ug/l  
 RT: 13.619 min Scan# 4336  
 Delta R.T. 0.000 min  
 Lab File: VN069237.D  
 Acq: 27 Oct 2021 23:55



Tgt Ion:119 Resp: 182928  
 Ion Ratio Lower Upper  
 119 100  
 134 25.8 12.9 38.6  
 91 25.4 12.8 38.4









#87  
1,3-Dichlorobenzene  
Concen: 20.228 ug/l  
RT: 13.619 min Scan# 4336  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

Tgt Ion:146 Resp: 90953  
Ion Ratio Lower Upper  
146 100  
111 42.7 21.4 64.2  
148 63.8 32.0 96.2

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Manual Integrations  
APPROVED


MMDadoda  
10/29/2021 1:25:47 PM



#88  
1,4-Dichlorobenzene  
Concen: 19.913 ug/l  
RT: 13.696 min Scan# 4365  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55


Tgt Ion:146 Resp: 88866  
Ion Ratio Lower Upper  
146 100  
111 42.1 20.6 61.8  
148 62.9 32.1 96.5





#89  
n-Butylbenzene  
Concen: 17.145 ug/l  
RT: 13.946 min Scan# 4458  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion: 91 Resp: 135146  
Ion Ratio Lower Upper  
91 100  
92 51.3 26.0 78.0  
134 24.4 12.5 37.5

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#90  
Hexachloroethane  
Concen: 16.518 ug/l  
RT: 14.211 min Scan# 4557  
Delta R.T. -0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion: 117 Resp: 32854  
Ion Ratio Lower Upper  
117 100  
201 75.4 37.1 111.5









#91  
1,2-Dichlorobenzene  
Concen: 20.945 ug/l  
RT: 13.991 min Scan# 4475  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55





Tgt Ion:146 Resp: 91485  
Ion Ratio Lower Upper  
146 100  
111 44.1 22.0 66.0  
148 64.0 32.1 96.5



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 23.443 ug/l  
RT: 14.603 min Scan# 4703  
Delta R.T. -0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Tgt Ion: 75 Resp: 15355  
Ion Ratio Lower Upper  
75 100  
155 73.8 38.8 116.4  
157 93.2 50.2 150.6



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06


Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#93  
1,2,4-Trichlorobenzene  
Concen: 17.452 ug/l  
RT: 15.268 min Scan# 4951  
Delta R.T. 0.003 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06




Tgt Ion:180 Resp: 37159  
Ion Ratio Lower Upper  
180 100  
182 93.8 47.9 143.7  
145 32.2 16.4 49.2

Manual Integrations  
APPROVED



MMDadoda  
10/29/2021 1:25:47 PM




#94  
Hexachlorobutadiene  
Concen: 17.749 ug/l  
RT: 15.373 min Scan# 4990  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55



Tgt Ion:225 Resp: 22075  
Ion Ratio Lower Upper  
225 100  
223 64.5 31.5 94.5  
227 60.3 31.6 94.7









#95  
Naphthalene  
Concen: 19.404 ug/l  
RT: 15.509 min Scan# 5041  
Delta R.T. 0.002 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 128   | Resp: | 110628 |
| Ion Ratio | Lower | Upper |        |
| 128       | 100   |       |        |
| 127       | 12.9  | 10.5  | 15.7   |
| 129       | 10.8  | 8.9   | 13.3   |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1027WBS06

Manual Integrations  
APPROVED

MMDadoda  
10/29/2021 1:25:47 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 18.613 ug/l  
RT: 15.702 min Scan# 5113  
Delta R.T. 0.000 min  
Lab File: VN069237.D  
Acq: 27 Oct 2021 23:55

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 180   | Resp: | 38258 |
| Ion Ratio | Lower | Upper |       |
| 180       | 100   |       |       |
| 182       | 93.8  | 48.1  | 144.3 |
| 145       | 34.0  | 17.4  | 52.2  |

