



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069176.D
 Acq On : 26 Oct 2021 8:41
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Title : SW846 8260
 Last Update : Wed Oct 13 05:44:40 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3988

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	16592	PASS
75	95	30	60	52.6	45760	PASS
95	95	100	100	100.0	86960	PASS
96	95	5	9	6.9	6035	PASS
173	174	0.00	2	1.0	625	PASS
174	95	50	100	73.7	64104	PASS
175	174	5	9	7.4	4744	PASS
176	174	95	101	98.4	63083	PASS
177	176	5	9	6.6	4147	PASS