

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
 Data File : VN075036.D
 Acq On : 27 Oct 2022 13:19
 Operator : JC\MD
 Sample : N5303-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1837

Quant Time: Oct 28 02:13:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	109318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	202910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	173771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	70262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	82410	55.950	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.900%
35) Dibromofluoromethane	8.171	113	63396	53.221	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.440%
50) Toluene-d8	10.571	98	245221	53.061	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.120%
62) 4-Bromofluorobenzene	12.853	95	72691	44.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.560%
Target Compounds						
						Qvalue
40) Benzene	8.612	78	2316	0.410	ug/l	99
52) Toluene	10.635	92	3860	1.023	ug/l	89
95) Naphthalene	15.647	128	3332	0.833	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
Data File : VN075036.D
Acq On : 27 Oct 2022 13:19
Operator : JC\MD
Sample : N5303-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1837

Quant Time: Oct 28 02:13:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

