

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
 Data File : VN075040.D
 Acq On : 27 Oct 2022 14:55
 Operator : JC\MD
 Sample : N5271-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-IDW-0002

Quant Time: Oct 28 02:14:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	122916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	226993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91863	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	100983	60.975	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.940%	
35) Dibromofluoromethane	8.171	113	77371	54.039	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.080%	
50) Toluene-d8	10.571	98	309455	55.708	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.420%	
62) 4-Bromofluorobenzene	12.853	95	98778	50.621	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.240%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	13617	19.418	ug/l	100
20) Methylene Chloride	5.289	84	51466	32.916	ug/l	97
37) Ethyl Acetate	7.571	43	8680	3.571	ug/l #	95
43) Isopropyl Acetate	8.694	43	71600	18.988	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
Data File : VN075040.D
Acq On : 27 Oct 2022 14:55
Operator : JC\MD
Sample : N5271-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-IDW-0002

Quant Time: Oct 28 02:14:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

