

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
 Data File : VN075041.D
 Acq On : 27 Oct 2022 15:19
 Operator : JC\MD
 Sample : N5273-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LIDEN-GAS-PIPELINE

Quant Time: Oct 28 02:14:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	120152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243451	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	222167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	87652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	100886	62.317	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.640%	
35) Dibromofluoromethane	8.171	113	76670	53.646	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.300%	
50) Toluene-d8	10.571	98	301803	54.430	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.860%	
62) 4-Bromofluorobenzene	12.853	95	96571	49.580	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	17809	25.980	ug/l	96
20) Methylene Chloride	5.283	84	4326706	2830.887	ug/l	94
30) Chloroform	7.971	83	2882	1.022	ug/l	97
37) Ethyl Acetate	7.571	43	7744	3.191	ug/l	97
43) Isopropyl Acetate	8.694	43	60199	15.994	ug/l #	69
95) Naphthalene	15.647	128	14057	2.817	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
Data File : VN075041.D
Acq On : 27 Oct 2022 15:19
Operator : JC\MD
Sample : N5273-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LIDEN-GAS-PIPELINE

Quant Time: Oct 28 02:14:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

