

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
 Data File : VN075043.D
 Acq On : 27 Oct 2022 16:07
 Operator : JC\MD
 Sample : N5284-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-3-WC

Quant Time: Oct 28 02:15:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	109737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	214092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	193657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	79550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	86280	58.353	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.700%
35) Dibromofluoromethane	8.171	113	65259	51.924	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.840%
50) Toluene-d8	10.571	98	267490	54.857	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.720%
62) 4-Bromofluorobenzene	12.853	95	84308	49.219	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.440%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	15876	25.358	ug/l	100
20) Methylene Chloride	5.283	84	44848	32.128	ug/l	97
30) Chloroform	7.971	83	4222	1.640	ug/l	93
37) Ethyl Acetate	7.571	43	12118	5.679	ug/l #	95
43) Isopropyl Acetate	8.688	43	156282	47.215	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
Data File : VN075043.D
Acq On : 27 Oct 2022 16:07
Operator : JC\MD
Sample : N5284-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-3-WC

Quant Time: Oct 28 02:15:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

