

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
 Data File : VN075046.D
 Acq On : 27 Oct 2022 17:19
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 28 02:15:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	101232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	186599	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	161476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	66641	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	76427	56.032	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.060%
35) Dibromofluoromethane	8.171	113	58980	53.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.680%
50) Toluene-d8	10.571	98	230516	54.239	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.480%
62) 4-Bromofluorobenzene	12.853	95	71030	47.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
Data File : VN075046.D
Acq On : 27 Oct 2022 17:19
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 28 02:15:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

