

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
 Data File : VN075049.D
 Acq On : 27 Oct 2022 18:31
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 28 02:15:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	101839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	189224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	160879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61615	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	77296	56.332	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.660%
35) Dibromofluoromethane	8.171	113	58514	52.675	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.360%
50) Toluene-d8	10.571	98	226903	52.649	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.300%
62) 4-Bromofluorobenzene	12.853	95	71443	47.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102722\
Data File : VN075049.D
Acq On : 27 Oct 2022 18:31
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 28 02:15:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

