

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102725\
 Data File : VN088111.D
 Acq On : 27 Oct 2025 14:13
 Operator : JC\MD
 Sample : Q3449-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CF-620-COMP-41

Quant Time: Oct 28 01:27:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	219087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	493422	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	481690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	227894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	212429	56.072	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.140%
35) Dibromofluoromethane	8.153	113	155362	46.869	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	10.547	98	576893	45.166	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.340%
62) 4-Bromofluorobenzene	12.829	95	219274	48.611	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.220%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	60858	31.975	ug/l	96
20) Methylene Chloride	5.265	84	42501	14.898	ug/l	89
43) Isopropyl Acetate	8.677	43	22140	2.307	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102725\
Data File : VN088111.D
Acq On : 27 Oct 2025 14:13
Operator : JC\MD
Sample : Q3449-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CF-620-COMP-41

Quant Time: Oct 28 01:27:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

