

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102725\
 Data File : VN088116.D
 Acq On : 27 Oct 2025 15:56
 Operator : JC\MD
 Sample : Q3451-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Oct 28 01:29:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	241150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	563402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	556355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	269948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	255482	61.267	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.540%	
35) Dibromofluoromethane	8.153	113	185549	49.023	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.040%	
50) Toluene-d8	10.547	98	672963	46.144	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.280%	
62) 4-Bromofluorobenzene	12.829	95	260924	50.659	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.320%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	41199	19.666	ug/l	94
20) Methylene Chloride	5.265	84	44480	14.203	ug/l	88
43) Isopropyl Acetate	8.677	43	31618	2.885	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102725\
Data File : VN088116.D
Acq On : 27 Oct 2025 15:56
Operator : JC\MD
Sample : Q3451-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Oct 28 01:29:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

