

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102725\
 Data File : VN088117.D
 Acq On : 27 Oct 2025 16:17
 Operator : JC\MD
 Sample : Q3452-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JB-1

Quant Time: Oct 28 01:29:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	242070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	547486	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	538103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	257845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	234463	56.013	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.020%
35) Dibromofluoromethane	8.147	113	173065	47.054	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	10.547	98	648356	45.749	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.500%
62) 4-Bromofluorobenzene	12.829	95	244033	48.757	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.520%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	39580	18.821	ug/l	99
20) Methylene Chloride	5.271	84	36860	11.860	ug/l	94
43) Isopropyl Acetate	8.677	43	28159	2.644	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102725\
Data File : VN088117.D
Acq On : 27 Oct 2025 16:17
Operator : JC\MD
Sample : Q3452-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JB-1

Quant Time: Oct 28 01:29:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

