

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102819\
 Data File : VN058985.D
 Acq On : 28 Oct 2019 10:31
 Operator : JC/SP
 Sample : VN1028MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1028MBL01

Quant Time: Oct 29 13:44:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	374998	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	595032	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	507470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	171972	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	271438	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	
35) Dibromofluoromethane	7.57	113	197054	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
50) Toluene-d8	10.08	98	753259	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
62) 4-Bromofluorobenzene	12.40	95	224668	40.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102819\
Data File : VN058985.D
Acq On : 28 Oct 2019 10:31
Operator : JC/SP
Sample : VN1028MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1028MBL01

Quant Time: Oct 29 13:44:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

