

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102819\
 Data File : VN058987.D
 Acq On : 28 Oct 2019 11:15
 Operator : JC/SP
 Sample : PB124239TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124239TB

Quant Time: Oct 29 13:46:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	374206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	592744	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	543641	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	188399	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	265702	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
35) Dibromofluoromethane	7.57	113	190460	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	10.08	98	770878	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.40	95	238230	42.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.76%	
Target Compounds						
16) Acetone	3.80	43	11906	4.390	ug/l	95
20) Methylene Chloride	4.53	84	12409	2.680	ug/l	96
43) Isopropyl Acetate	8.15	43	129484	12.210	ug/l #	91
80) 1,3,5-Trimethylbenzene	12.73	105	8493	0.734	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102819\
Data File : VN058987.D
Acq On : 28 Oct 2019 11:15
Operator : JC/SP
Sample : PB124239TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124239TB

Quant Time: Oct 29 13:46:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

