

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102819\
 Data File : VN058992.D
 Acq On : 28 Oct 2019 13:06
 Operator : JC/SP
 Sample : K5564-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CSA-2-1-1

Quant Time: Nov 07 04:49:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	355569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	565343	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	513871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	185669	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	247417	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
35) Dibromofluoromethane	7.57	113	181265	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	10.08	98	725527	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	12.41	95	227792	42.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.98%	
Target Compounds						
16) Acetone	3.80	43	17134	6.649	ug/l	100
20) Methylene Chloride	4.53	84	14638	3.327	ug/l	97
43) Isopropyl Acetate	8.15	43	123488	12.209	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102819\
Data File : VN058992.D
Acq On : 28 Oct 2019 13:06
Operator : JC/SP
Sample : K5564-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CSA-2-1-1

Quant Time: Nov 07 04:49:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

