

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102819\
 Data File : VN058994.D
 Acq On : 28 Oct 2019 14:42
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 29 14:02:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.623	0.548	12.0	83	0.00
3 P	Chloromethane	0.772	0.662	14.2	83	0.00
4 C	Vinyl Chloride	0.767	0.690	10.0#	86	0.00
5 T	Bromomethane	0.480	0.417	13.1	83	0.00
6 T	Chloroethane	0.492	0.437	11.2	87	0.00
7 T	Trichlorofluoromethane	0.949	0.898	5.4	89	0.00
8 T	Diethyl Ether	0.328	0.329	-0.3	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.526	0.504	4.2	94	0.00
10 T	Methyl Iodide	0.692	0.627	9.4	86	0.00
11 T	Tert butyl alcohol	0.122	0.107	12.3	86	0.00
12 CM	1,1-Dichloroethene	0.493	0.474	3.9#	92	0.00
13 T	Acrolein	0.084	0.064	23.8	71	0.00
14 T	Allyl chloride	0.997	0.910	8.7	86	0.00
15 T	Acrylonitrile	0.306	0.301	1.6	90	0.00
16 T	Acetone	0.362	0.337	6.9	84	0.00
17 T	Carbon Disulfide	1.588	1.361	14.3	82	0.00
18 T	Methyl Acetate	0.824	0.798	3.2	97	0.00
19 T	Methyl tert-butyl Ether	1.726	1.798	-4.2	96	0.00
20 T	Methylene Chloride	0.619	0.584	5.7	96	0.00
21 T	trans-1,2-Dichloroethene	0.544	0.528	2.9	92	0.00
22 T	Diisopropyl ether	1.951	2.107	-8.0	99	0.00
23 T	Vinyl Acetate	1.567	1.716	-9.5	93	0.00
24 P	1,1-Dichloroethane	1.131	1.112	1.7	95	0.00
25 T	2-Butanone	0.470	0.442	6.0	85	0.00
26 T	2,2-Dichloropropane	0.978	0.985	-0.7	96	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.632	-1.6	95	0.00
28 T	Bromochloromethane	0.377	0.369	2.1	93	0.00
29 T	Tetrahydrofuran	0.288	0.278	3.5	87	0.00
30 C	Chloroform	1.135	1.105	2.6#	95	0.00
31 T	Cyclohexane	1.054	0.982	6.8	90	0.00
32 T	1,1,1-Trichloroethane	0.956	0.952	0.4	94	0.00
33 S	1,2-Dichloroethane-d4	0.733	0.701	4.4	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.329	0.339	-3.0	95	0.00
36 T	1,1-Dichloropropene	0.512	0.527	-2.9	92	0.00
37 T	Ethyl Acetate	0.558	0.554	0.7	87	0.00
38 T	Carbon Tetrachloride	0.527	0.523	0.8	92	0.00
39 T	Methylcyclohexane	0.575	0.606	-5.4	93	0.00
40 TM	Benzene	1.503	1.563	-4.0	94	0.00
41 T	Methacrylonitrile	0.248	0.285	-14.9	109	0.00
42 TM	1,2-Dichloroethane	0.579	0.576	0.5	92	0.00
43 T	Isopropyl Acetate	0.895	0.920	-2.8	92	0.00
44 TM	Trichloroethene	0.366	0.377	-3.0	95	0.00
45 C	1,2-Dichloropropane	0.412	0.439	-6.6#	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102819\
 Data File : VN058994.D
 Acq On : 28 Oct 2019 14:42
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 29 14:02:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.261	0.267	-2.3	95	0.00
47 T	Bromodichloromethane	0.537	0.565	-5.2	96	0.00
48 T	Methyl methacrylate	0.406	0.425	-4.7	90	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	92	0.00
50 S	Toluene-d8	1.276	1.289	-1.0	93	0.00
51 T	4-Methyl-2-Pentanone	0.522	0.560	-7.3	89	0.00
52 CM	Toluene	0.891	0.963	-8.1#	95	0.00
53 T	t-1,3-Dichloropropene	0.595	0.643	-8.1	96	0.00
54 T	cis-1,3-Dichloropropene	0.635	0.693	-9.1	96	0.00
55 T	1,1,2-Trichloroethane	0.367	0.382	-4.1	97	0.00
56 T	Ethyl methacrylate	0.540	0.605	-12.0	95	0.00
57 T	1,3-Dichloropropane	0.645	0.676	-4.8	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.260	-11.1	94	0.00
59 T	2-Hexanone	0.404	0.424	-5.0	87	0.00
60 T	Dibromochloromethane	0.395	0.428	-8.4	98	0.00
61 T	1,2-Dibromoethane	0.376	0.390	-3.7	93	0.00
62 S	4-Bromofluorobenzene	0.469	0.487	-3.8	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.350	0.354	-1.1	93	0.00
65 PM	Chlorobenzene	1.018	1.089	-7.0	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.416	-8.3	98	0.00
67 C	Ethyl Benzene	1.786	2.006	-12.3#	95	0.00
68 T	m/p-Xylenes	0.662	0.746	-12.7	95	0.00
69 T	o-Xylene	0.632	0.712	-12.7	96	0.00
70 T	Styrene	1.030	1.198	-16.3	96	0.00
71 P	Bromoform	0.298	0.322	-8.1	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.654	3.900	-6.7	96	0.00
74 T	N-amyl acetate	1.657	1.734	-4.6	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.274	0.9	94	0.00
76 T	1,2,3-Trichloropropane	1.159	1.248	-7.7	106	0.00
77 T	Bromobenzene	0.901	0.935	-3.8	98	0.00
78 T	n-propylbenzene	4.319	4.695	-8.7	96	0.00
79 T	2-Chlorotoluene	2.644	2.743	-3.7	96	0.00
80 T	1,3,5-Trimethylbenzene	3.070	3.365	-9.6	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.457	-2.7	94	0.00
82 T	4-Chlorotoluene	2.702	2.872	-6.3	96	0.00
83 T	tert-Butylbenzene	2.625	2.853	-8.7	97	0.00
84 T	1,2,4-Trimethylbenzene	3.028	3.357	-10.9	95	0.00
85 T	sec-Butylbenzene	3.500	3.868	-10.5	96	0.00
86 T	p-Isopropyltoluene	3.088	3.521	-14.0	97	0.00
87 T	1,3-Dichlorobenzene	1.610	1.711	-6.3	98	0.00
88 T	1,4-Dichlorobenzene	1.612	1.684	-4.5	96	0.00
89 T	n-Butylbenzene	2.837	3.228	-13.8	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102819\
Data File : VN058994.D
Acq On : 28 Oct 2019 14:42
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 29 14:02:22 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.644	-2.9	97	0.00
91 T	1,2-Dichlorobenzene	1.583	1.674	-5.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.291	0.254	12.7	84	0.00
93 T	1,2,4-Trichlorobenzene	0.873	0.992	-13.6	97	0.00
94 T	Hexachlorobutadiene	0.513	0.533	-3.9	97	0.00
95 T	Naphthalene	2.436	2.622	-7.6	88	0.00
96 T	1,2,3-Trichlorobenzene	0.858	0.937	-9.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6