


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102819\
Data File : VN058983.D
Acq On : 28 Oct 2019 8:30
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Title : SW846 8260
Last Update : Sat Oct 19 00:16:20 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	15410	PASS
75	95	30	60	53.5	40015	PASS
95	95	100	100	100.0	74813	PASS
96	95	5	9	7.2	5402	PASS
173	174	0.00	2	1.2	691	PASS
174	95	50	100	76.3	57077	PASS
175	174	5	9	7.8	4458	PASS
176	174	95	101	98.2	56077	PASS
177	176	5	9	6.8	3834	PASS