

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102820\
 Data File : VN064361.D
 Acq On : 28 Oct 2020 18:43
 Operator : JC/MD
 Sample : L4528-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHRT-27339

Manual Integrations
 APPROVED

MMDadoda
 10/29/2020 12:01:42 PM

Quant Time: Oct 29 03:52:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	247212	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	438258	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	432344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178681	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	196955	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
35) Dibromofluoromethane	7.55	113	144508	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	10.06	98	540358	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	12.38	95	233308	55.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.70%	
Target Compounds						
16) Acetone	3.78	43	16653	5.099	ug/l	95
18) Methyl Acetate	4.27	43	4493m	1.506	ug/l	
20) Methylene Chloride	4.50	84	6008	2.132	ug/l	90
43) Isopropyl Acetate	8.13	43	91344	11.960	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102820\
Data File : VN064361.D
Acq On : 28 Oct 2020 18:43
Operator : JC/MD
Sample : L4528-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHRT-27339

Manual Integrations
APPROVED

MMDadoda
10/29/2020 12:01:42 PM

Quant Time: Oct 29 03:52:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

