

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069266.D
 Acq On : 28 Oct 2021 23:10
 Operator : JC/MD
 Sample : VN1028WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1028WBL02

Quant Time: Oct 29 07:54:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	314014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	549621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	535141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	199246	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	209771	46.277	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.560%
35) Dibromofluoromethane	8.024	113	161505	47.139	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.280%
50) Toluene-d8	10.443	98	685665	50.053	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.100%
62) 4-Bromofluorobenzene	12.733	95	249472	48.261	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069266.D
Acq On : 28 Oct 2021 23:10
Operator : JC/MD
Sample : VN1028WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1028WBL02

Quant Time: Oct 29 07:54:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

