

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069270.D
 Acq On : 29 Oct 2021 00:53
 Operator : JC/MD
 Sample : M4338-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DDC-4-PS

Quant Time: Oct 29 07:55:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	313693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	556141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	538225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	200580	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	210541	46.495	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.980%
35) Dibromofluoromethane	8.024	113	158547	45.733	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.460%
50) Toluene-d8	10.443	98	689507	49.743	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.480%
62) 4-Bromofluorobenzene	12.733	95	246074	47.045	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.100%
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.979	164	21349	6.210	ug/l	99
90) Hexachloroethane	14.211	117	3456	3.272	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069270.D
Acq On : 29 Oct 2021 00:53
Operator : JC/MD
Sample : M4338-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DDC-4-PS

Quant Time: Oct 29 07:55:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

