

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069271.D
 Acq On : 29 Oct 2021 1:19
 Operator : JC/MD
 Sample : M4338-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FD-01

Quant Time: Oct 29 07:56:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	307067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	540674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	525311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	194334	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	205496	46.360	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.720%
35) Dibromofluoromethane	8.024	113	157137	46.623	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.240%
50) Toluene-d8	10.443	98	673688	49.993	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.980%
62) 4-Bromofluorobenzene	12.733	95	241137	47.420	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069271.D
Acq On : 29 Oct 2021 1:19
Operator : JC/MD
Sample : M4338-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FD-01

Quant Time: Oct 29 07:56:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

