

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069275.D
 Acq On : 29 Oct 2021 3:01
 Operator : JC/MD
 Sample : M4338-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-15D

Quant Time: Oct 29 07:56:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	301983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	539150	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	521024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	195117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	205900	47.233	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.460%
35) Dibromofluoromethane	8.024	113	155795	46.355	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.720%
50) Toluene-d8	10.443	98	672394	50.038	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.080%
62) 4-Bromofluorobenzene	12.733	95	242453	47.814	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.620%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	5281	2.843	ug/l	98
30) Chloroform	7.820	83	3366	0.539	ug/l	99
64) Tetrachloroethene	10.977	164	11675	3.508	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069275.D
Acq On : 29 Oct 2021 3:01
Operator : JC/MD
Sample : M4338-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-15D

Quant Time: Oct 29 07:56:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

