

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069280.D
 Acq On : 29 Oct 2021 5:10
 Operator : JC/MD
 Sample : PB140292ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140292ZHE#01

Quant Time: Oct 29 07:57:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	302918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	542221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	550651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	206310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	205983	47.106	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.220%		
35) Dibromofluoromethane	8.024	113	155532	46.015	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.040%		
50) Toluene-d8	10.443	98	684618	50.659	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.320%		
62) 4-Bromofluorobenzene	12.734	95	249198	48.865	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.740%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	12624	6.776	ug/l	100
20) Methylene Chloride	5.098	84	12590	3.523	ug/l	92
43) Isopropyl Acetate	8.566	43	29172	3.570	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069280.D
Acq On : 29 Oct 2021 5:10
Operator : JC/MD
Sample : PB140292ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140292ZHE#01

Quant Time: Oct 29 07:57:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

