

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069281.D
 Acq On : 29 Oct 2021 5:35
 Operator : JC/MD
 Sample : PB140292ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140292ZHE#02

Quant Time: Oct 29 07:57:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	285515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	512147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	507973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	186004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	193789	47.019	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.040%
35) Dibromofluoromethane	8.024	113	147459	46.188	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.380%
50) Toluene-d8	10.443	98	642521	50.336	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.734	95	231034	47.964	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.920%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	11954	6.807	ug/l	95
18) Methyl Acetate	4.867	43	5982	1.002	ug/l	93
20) Methylene Chloride	5.095	84	12322	3.658	ug/l	95
43) Isopropyl Acetate	8.568	43	34189	4.319	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069281.D
Acq On : 29 Oct 2021 5:35
Operator : JC/MD
Sample : PB140292ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140292ZHE#02

Quant Time: Oct 29 07:57:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

