

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
 Data File : VN069284.D
 Acq On : 29 Oct 2021 6:52
 Operator : JC/MD
 Sample : M4378-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WASTE-CHAR-(EXCESS)

Quant Time: Oct 29 07:58:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	298715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	535312	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	525006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	202786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	201586	46.749	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.500%		
35) Dibromofluoromethane	8.024	113	153150	45.895	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.800%		
50) Toluene-d8	10.443	98	664927	49.837	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.680%		
62) 4-Bromofluorobenzene	12.734	95	245084	48.679	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.360%		
Target Compounds						
16) Acetone	4.296	43	53604	29.177	ug/l	99
20) Methylene Chloride	5.093	84	15405	4.372	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102821\
Data File : VN069284.D
Acq On : 29 Oct 2021 6:52
Operator : JC/MD
Sample : M4378-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WASTE-CHAR-(EXCESS)

Quant Time: Oct 29 07:58:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

