

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102822\
 Data File : VN075068.D
 Acq On : 28 Oct 2022 18:12
 Operator : JC\MD
 Sample : N5313-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMPOSITE-NW-QUARTER

Quant Time: Oct 31 03:09:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	105113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	200523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	185467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	71903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	83705	59.102	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.200%
35) Dibromofluoromethane	8.177	113	63996	54.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.720%
50) Toluene-d8	10.571	98	249933	54.725	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.440%
62) 4-Bromofluorobenzene	12.853	95	78658	49.028	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.060%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	10134	16.899	ug/l	97
20) Methylene Chloride	5.283	84	59044	44.159	ug/l	95
37) Ethyl Acetate	7.571	43	16341	8.176	ug/l	98
43) Isopropyl Acetate	8.694	43	70203	22.645	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102822\
Data File : VN075068.D
Acq On : 28 Oct 2022 18:12
Operator : JC\MD
Sample : N5313-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMPOSITE-NW-QUARTER

Quant Time: Oct 31 03:09:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

