

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102825\
 Data File : VN088129.D
 Acq On : 28 Oct 2025 11:17
 Operator : JC\MD
 Sample : VN1028WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1028WBL01

Quant Time: Oct 29 02:18:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	231261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	516618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	484454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	238378	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	219109	54.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.580%
35) Dibromofluoromethane	8.153	113	162498	46.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.640%
50) Toluene-d8	10.547	98	598983	44.790	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.580%
62) 4-Bromofluorobenzene	12.829	95	227915	48.258	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102825\
Data File : VN088129.D
Acq On : 28 Oct 2025 11:17
Operator : JC\MD
Sample : VN1028WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1028WBL01

Quant Time: Oct 29 02:18:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

