

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102825\
 Data File : VN088143.D
 Acq On : 28 Oct 2025 16:24
 Operator : JC\MD
 Sample : Q3477-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOUTH-MAHWAH-WATER

Quant Time: Oct 29 02:28:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	239370	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	541511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	502020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	242841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	221126	53.422	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.840%
35) Dibromofluoromethane	8.147	113	163794	45.024	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.040%
50) Toluene-d8	10.547	98	612015	43.661	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.320%
62) 4-Bromofluorobenzene	12.829	95	233958	47.260	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.520%
Target Compounds						
16) Acetone	4.424	43	9724	4.676	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102825\
Data File : VN088143.D
Acq On : 28 Oct 2025 16:24
Operator : JC\MD
Sample : Q3477-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOUTH-MAHWAH-WATER

Quant Time: Oct 29 02:28:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

