

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102919\
 Data File : VN058999.D
 Acq On : 29 Oct 2019 10:59
 Operator : JC/SP
 Sample : PB124371TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 30 03:53:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	441613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	702396	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	642717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	233778	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	307164	47.42	ug/l	0.00
Spiked Amount			Recovery	=	94.84%	
35) Dibromofluoromethane	7.57	113	224765	48.67	ug/l	0.00
Spiked Amount			Recovery	=	97.34%	
50) Toluene-d8	10.08	98	892765	49.80	ug/l	0.00
Spiked Amount			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	12.40	95	287138	43.62	ug/l	0.00
Spiked Amount			Recovery	=	87.24%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.76	59	349	0.325	ug/l #	72
16) Acetone	3.80	43	16535	5.166	ug/l	96
18) Methyl Acetate	4.30	43	7516	1.033	ug/l #	83
20) Methylene Chloride	4.53	84	15234	2.788	ug/l	93
25) 2-Butanone	6.81	43	1036	0.249	ug/l #	51
31) Cyclohexane	7.65	56	10380	1.115	ug/l #	13
36) 1,1-Dichloropropene	7.65	75	36294	5.047	ug/l #	50
38) Carbon Tetrachloride	7.65	117	44675	6.030	ug/l #	17
43) Isopropyl Acetate	8.15	43	130142	10.356	ug/l #	91
45) 1,2-Dichloropropane	9.26	63	2967	0.512	ug/l #	44
48) Methyl methacrylate	9.17	41	7770	1.363	ug/l #	14
49) 1,4-Dioxane	9.26	88	192	2.051	ug/l #	1
51) 4-Methyl-2-Pentanone	9.98	43	8586	1.170	ug/l	90
54) cis-1,3-Dichloropropene	9.90	75	49929	5.595	ug/l #	64
56) Ethyl methacrylate	10.54	69	2435	2.367	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	9309	1.028	ug/l #	44
59) 2-Hexanone	10.76	43	132691	23.395	ug/l #	53
70) Styrene	11.96	104	160	1.724	ug/l #	69
71) Bromoform	12.41	173	1643	0.429	ug/l #	1
74) N-amyl acetate	12.04	43	5712	0.737	ug/l #	49
76) 1,2,3-Trichloropropane	12.40	75	148973	27.496	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	148973	71.628	ug/l #	10
93) 1,2,4-Trichlorobenzene	14.91	180	410	2.604	ug/l #	44
95) Naphthalene	15.14	128	1794	2.353	ug/l #	89
96) 1,2,3-Trichlorobenzene	15.30	180	629	2.526	ug/l	70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102919\
Data File : VN058999.D
Acq On : 29 Oct 2019 10:59
Operator : JC/SP
Sample : PB124371TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Oct 30 03:53:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1827

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 441613

Ion Ratio Lower Upper

168 100

99 59.1 48.1 72.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.65

200000

150000

100000



50000

0

7.60

7.70

Time-->



#11

Tert butyl alcohol

Concen: 0.325 ug/l

RT: 4.76 min Scan# 928

Delta R.T. 0.00 min

Lab File: VN058999.D


Acq: 29 Oct 2019 10:59

Tgt Ion: 59 Resp: 349

Ion Ratio Lower Upper

59 100

57 0.0 8.6 12.8#



Abundance Ion 59.00 (58.70 to 59.70): VN

Ion 57.00 (56.70 to 57.70): VN

4.76

400

300

200

100


0

4.74

4.76

4.78

Time-->









#16
Acetone
Concen: 5.166 ug/l
RT: 3.80 min Scan# 629
Delta R.T. 0.01 min
Lab File: VN058999.D
Acq: 29 Oct 2019 10:59




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 16535
Ion Ratio Lower Upper
43 100
58 31.1 23.1 34.7



#18
Methyl Acetate
Concen: 1.033 ug/l
RT: 4.30 min Scan# 787
Delta R.T. 0.01 min
Lab File: VN058999.D
Acq: 29 Oct 2019 10:59

Tgt Ion: 43 Resp: 7516
Ion Ratio Lower Upper
43 100
74 13.3 17.1 25.7#









#20
Methvlene Chloride
Concen: 2.788 ug/l
RT: 4.53 min Scan# 858
Delta R.T. 0.00 min
Lab File: VN058999.D
Acq: 29 Oct 2019 10:59

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 84 Resp: 15234
Ion Ratio Lower Upper
84 100
49 124.7 109.6 164.4
51 42.2 33.8 50.6
86 67.4 51.5 77.3




Abundance Ion 83.90 (83.60 to 84.60): VN0
Ion 48.90 (48.60 to 49.60): VN0
Ion 50.90 (50.60 to 51.60): VN0
Ion 85.90 (85.60 to 86.60): VN0






#25
2-Butanone
Concen: 0.249 ug/l
RT: 6.81 min Scan# 1566
Delta R.T. -0.00 min
Lab File: VN058999.D
Acq: 29 Oct 2019 10:59

Tgt Ion: 43 Resp: 1036
Ion Ratio Lower Upper
43 100
72 0.0 19.1 28.7#



Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 72.00 (71.70 to 72.70): VN0





#31

Cyclohexane

Concen: 1.115 ug/l

RT: 7.65 min Scan# 1827

Delta R.T. 0.01 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 56 Resp: 10380

Ion Ratio Lower Upper




56 100

69 159.1 26.2 39.4#

84 111.3 65.6 98.4#



Abundance Ion 56.00 (55.70 to 56.70): VNC
Ion 69.00 (68.70 to 69.70): VNC
Ion 84.00 (83.70 to 84.70): VNC



#33

1,2-Dichloroethane-d4

Concen: 47.424 ug/l

RT: 8.01 min Scan# 1939

Delta R.T. 0.00 min

Lab File: VN058999.D


Acq: 29 Oct 2019 10:59

Tgt Ion: 65 Resp: 307164



Ion Ratio Lower Upper


65 100

67 52.4 0.0 103.2



Abundance Ion 64.90 (64.60 to 65.60): VNC
Ion 66.90 (66.60 to 67.60): VNC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2114

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:114 Resp: 702396

Ion Ratio Lower Upper




114 100

63 22.5 0.0 44.4

88 16.3 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 48.671 ug/l

RT: 7.57 min Scan# 1803

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59


Tgt Ion:113 Resp: 224765

Ion Ratio Lower Upper



113 100


111 102.8 82.0 123.0

192 19.1 15.0 22.6



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 5.047 ug/l
 RT: 7.65 min Scan# 1827
 Delta R.T. -0.13 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
75	100		
110	8.2	16.3	48.9#
77	0.0	24.6	36.8#



#38
 Carbon Tetrachloride
 Concen: 6.030 ug/l
 RT: 7.65 min Scan# 1827
 Delta R.T. -0.11 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59

Tgt Ion	Ratio	Lower	Upper
117	100		
119	5.4	75.7	113.5#
121	0.0	25.0	37.6#





#43

Isopropyl Acetate

Concen: 10.356 ug/l

RT: 8.15 min Scan# 1982

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 43 Resp: 130142

Ion Ratio Lower Upper

43 100

61 18.3 20.2 30.2#




87 11.8 9.5 14.3



Abundance Ion 43.00 (42.70 to 43.70): VN058999.D (-2188) (-)

Ion 61.00 (60.70 to 61.70): VN058999.D (-2188) (-)

Ion 87.00 (86.70 to 87.70): VN058999.D (-2188) (-)



#45

1,2-Dichloropropane

Concen: 0.512 ug/l

RT: 9.26 min Scan# 2327

Delta R.T. 0.15 min

Lab File: VN058999.D


Acq: 29 Oct 2019 10:59

Tgt Ion: 63 Resp: 2967

Ion Ratio Lower Upper



63 100


65 0.0 24.5 36.7#



Abundance Ion 62.90 (62.60 to 63.60): VN058999.D (-2188) (-)

Ion 64.90 (64.60 to 65.60): VN058999.D (-2188) (-)









#48
Methvl methacrylate
Concen: 1.363 ug/l
RT: 9.17 min Scan# 2299
Delta R.T. -0.02 min
Lab File: VN058999.D
Acq: 29 Oct 2019 10:59




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 41 Resp: 7770
Ion Ratio Lower Upper
41 100
69 0.0 63.8 95.6#
39 0.0 47.8 71.8#



#49
1,4-Dioxane
Concen: 2.051 ug/l
RT: 9.26 min Scan# 2327
Delta R.T. 0.07 min
Lab File: VN058999.D
Acq: 29 Oct 2019 10:59

Tgt Ion: 88 Resp: 192
Ion Ratio Lower Upper
88 100
43 1182802.6 29.1 43.7#
58 5170.3 59.4 89.0#





#50

Toluene-d8

Concen: 49.801 ug/l

RT: 10.08 min Scan# 2584

Delta R.T. -0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 98 Resp: 892765




Ion Ratio Lower Upper

98 100

100 64.1 51.4 77.2



Abundance Ion 98.00 (97.70 to 98.70): VN058846.D (-2568) (-)



#51

4-Methyl-2-Pentanone

Concen: 1.170 ug/l

RT: 9.98 min Scan# 2551

Delta R.T. 0.00 min

Lab File: VN058999.D


Acq: 29 Oct 2019 10:59

Tgt Ion: 43 Resp: 8586



Ion Ratio Lower Upper


43 100

58 31.5 29.9 44.9



Abundance Ion 43.00 (42.70 to 43.70): VN058846.D (-2533) (-)









#54

cis-1,3-Dichloropropene
 Concen: 5.595 ug/l
 RT: 9.90 min Scan# 2526
 Delta R.T. 0.06 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59

Instrument :
 MSVOA_N
 ClientSampleId :




Tot Ion	Ratio	Lower	Upper
75	100		
77	0.7	24.7	37.1#
39	35.8	43.5	65.3#




#56

Ethyl methacrylate
 Concen: 2.367 ug/l
 RT: 10.54 min Scan# 2726
 Delta R.T. 0.12 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59

Tgt Ion	Ratio	Lower	Upper
69	100		
41	4459.3	57.2	85.8#
39	1850.7	31.0	46.4#





#57

1,3-Dichloropropane

Concen: 1.028 ug/l

RT: 10.76 min Scan# 2794

Delta R.T. 0.05 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 76 Resp: 9309

Ion Ratio Lower Upper

76 100

78 0.7 25.4 38.0#



#59

2-Hexanone

Concen: 23.395 ug/l

RT: 10.76 min Scan# 2795

Delta R.T. 0.01 min

Lab File: VN058999.D




Acq: 29 Oct 2019 10:59


Tgt Ion: 43 Resp: 132691

Ion Ratio Lower Upper

43 100

58 18.7 25.9 77.7#





#62

4-Bromofluorobenzene

Concen: 43.617 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 95 Resp: 287138

Ion Ratio Lower Upper

95 100

174 76.3 0.0 147.6




176 74.5 0.0 143.2



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59


Tgt Ion: 117 Resp: 642717

Ion Ratio Lower Upper

117 100

82 58.2 46.6 69.8



119 31.9 25.6 38.4




Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0





#70

Stvrene

Concen: 1.724 ug/l

RT: 11.96 min Scan# 3169

Delta R.T. -0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:104 Resp: 160

Ion Ratio Lower Upper

104 100

78 18.8 42.8 64.2#




103 65.0 44.3 66.5



Abundance Ion 104.00 (103.70 to 104.70): V

Ion 78.00 (77.70 to 78.70): VN

Ion 103.00 (102.70 to 103.70): V



#71

Bromoform

Concen: 0.429 ug/l

RT: 12.41 min Scan# 3307

Delta R.T. 0.28 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59


Tgt Ion:173 Resp: 1643

Ion Ratio Lower Upper

173 100

175 775.3 24.7 74.1#



254 0.0 0.0 0.0




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3600

Delta R.T. 0.00 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:152 Resp: 233778

Ion Ratio Lower Upper

152 100

115 60.5 30.5 91.5




150 159.8 0.0 344.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.737 ug/l

RT: 12.04 min Scan# 3192

Delta R.T. -0.04 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Tgt Ion: 43 Resp: 5712


Ion Ratio Lower Upper

43 100

70 0.0 30.9 46.3#

55 6.5 19.7 29.5#

61 0.0 18.2 27.4#






Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 70.00 (69.70 to 70.70): VNO

Ion 55.00 (54.70 to 55.70): VNO

Ion 61.00 (60.70 to 61.70): VNO





#76

1,2,3-Trichloropropane

Concen: 27.496 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 75 Resp: 148973

Ion Ratio Lower Upper




75 100

77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#81

trans-1,4-Dichloro-2-butene

Concen: 71.628 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN058999.D

Acq: 29 Oct 2019 10:59


Tgt Ion: 75 Resp: 148973

Ion Ratio Lower Upper

75 100

53 0.0 78.6 118.0#



89 0.0 35.4 53.0#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0









#93
 1,2,4-Trichlorobenzene
 Concen: 2.604 ug/l
 RT: 14.91 min Scan# 4086
 Delta R.T. -0.01 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion:180 Resp: 410
 Ion Ratio Lower Upper
 180 100
 182 40.5 47.3 142.0#
 145 0.0 15.3 45.8#




Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.80 (144.50 to 145.50): V






#95
 Naphthalene
 Concen: 2.353 ug/l
 RT: 15.14 min Scan# 4156
 Delta R.T. 0.01 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59

Tgt Ion:128 Resp: 1794
 Ion Ratio Lower Upper
 128 100
 127 8.4 10.2 15.2#
 129 6.0 8.4 12.6#



Abundance Ion 128.00 (127.70 to 128.70): V
 Ion 127.00 (126.70 to 127.70): V
 Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 2.526 ug/l
 RT: 15.30 min Scan# 4207
 Delta R.T. 0.00 min
 Lab File: VN058999.D
 Acq: 29 Oct 2019 10:59

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
180	100		
182	58.7	47.9	143.7
145	27.7	15.9	47.7

