

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102919\
 Data File : VN058999.D
 Acq On : 29 Oct 2019 10:59
 Operator : JC/SP
 Sample : PB124371TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124371TB

Quant Time: Oct 30 03:21:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	441613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	702396	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	642717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	233778	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	307164	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
35) Dibromofluoromethane	7.57	113	224765	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	10.08	98	892765	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	12.40	95	287138	43.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.24%	
Target Compounds						
16) Acetone	3.80	43	16535	5.166	ug/l	96
18) Methyl Acetate	4.30	43	7516	1.033	ug/l #	83
20) Methylene Chloride	4.53	84	15234	2.788	ug/l	93
43) Isopropyl Acetate	8.15	43	130142	10.356	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102919\
Data File : VN058999.D
Acq On : 29 Oct 2019 10:59
Operator : JC/SP
Sample : PB124371TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124371TB

Quant Time: Oct 30 03:21:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

