

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102919\
 Data File : VN059001.D
 Acq On : 29 Oct 2019 11:43
 Operator : JC/SP
 Sample : K5587-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CSA-2-1-9

Quant Time: Oct 30 03:35:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	424837	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	661547	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	591088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	209928	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	283759	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
35) Dibromofluoromethane	7.57	113	210786	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	10.08	98	832151	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	12.40	95	261285	42.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.28%	
Target Compounds						
20) Methylene Chloride	4.53	84	7712	1.467	ug/l	89
43) Isopropyl Acetate	8.15	43	99171	8.379	ug/l #	92
95) Naphthalene	15.14	128	9300	2.964	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102919\
Data File : VN059001.D
Acq On : 29 Oct 2019 11:43
Operator : JC/SP
Sample : K5587-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CSA-2-1-9

Quant Time: Oct 30 03:35:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

