

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102919\
 Data File : VN059011.D
 Acq On : 29 Oct 2019 15:23
 Operator : JC/SP
 Sample : K5576-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CSA-2-1-3

Quant Time: Oct 30 04:28:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	384308	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	603805	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	554673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	197307	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	270936	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
35) Dibromofluoromethane	7.57	113	196054	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	10.08	98	776634	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	12.40	95	243558	43.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.08%	
Target Compounds						
16) Acetone	3.80	43	16902	6.068	ug/l	97
18) Methyl Acetate	4.31	43	10114	1.597	ug/l #	81
20) Methylene Chloride	4.53	84	7614	1.601	ug/l	93
43) Isopropyl Acetate	8.15	43	72150	6.679	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102919\
Data File : VN059011.D
Acq On : 29 Oct 2019 15:23
Operator : JC/SP
Sample : K5576-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CSA-2-1-3

Quant Time: Oct 30 04:28:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

