

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102920\
 Data File : VN064370.D
 Acq On : 29 Oct 2020 13:47
 Operator : JC/MD
 Sample : L4554-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Oct 30 07:29:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	406858	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	416552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	190470	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	185358	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
35) Dibromofluoromethane	7.55	113	135409	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
50) Toluene-d8	10.06	98	507462	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.38	95	224494	57.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.74%	
Target Compounds						
					Qvalue	
31) Cyclohexane	7.61	56	22756	5.169	ug/l #	82
39) Methylcyclohexane	9.05	83	20565	5.051	ug/l #	89
40) Benzene	8.00	78	41327	3.616	ug/l	96
73) Isopropylbenzene	12.22	105	140720	9.864	ug/l	97
78) n-propylbenzene	12.57	91	225154	13.941	ug/l	99
83) tert-Butylbenzene	12.97	119	3581	0.362	ug/l	86
85) sec-Butylbenzene	13.14	105	70674	5.514	ug/l	88
89) n-Butylbenzene	13.59	91	40049	3.937	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102920\
Data File : VN064370.D
Acq On : 29 Oct 2020 13:47
Operator : JC/MD
Sample : L4554-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Oct 30 07:29:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

