

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102920\
 Data File : VN064372.D
 Acq On : 29 Oct 2020 14:34
 Operator : JC/MD
 Sample : L4554-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: Oct 30 07:34:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	246058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	432046	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	421927	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	193585	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	194676	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
35) Dibromofluoromethane	7.55	113	142307	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	10.06	98	523193	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	12.38	95	224677	54.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.14%	
Target Compounds						
73) Isopropylbenzene	12.23	105	10848	0.748	ug/l	99
78) n-propylbenzene	12.57	91	5341	0.325	ug/l	94
83) tert-Butylbenzene	12.97	119	5152	0.512	ug/l	90
85) sec-Butylbenzene	13.14	105	42627	3.272	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102920\
Data File : VN064372.D
Acq On : 29 Oct 2020 14:34
Operator : JC/MD
Sample : L4554-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: Oct 30 07:34:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

