

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102920\
 Data File : VN064375.D
 Acq On : 29 Oct 2020 15:45
 Operator : JC/MD
 Sample : L4227-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-02-102820

Quant Time: Oct 30 07:43:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	248615	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	439958	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	432688	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	190108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	196721	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
35) Dibromofluoromethane	7.55	113	145740	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	10.06	98	537736	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	12.38	95	234987	55.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.08%	
Target Compounds						
20) Methylene Chloride	4.51	84	13229	4.669	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102920\
Data File : VN064375.D
Acq On : 29 Oct 2020 15:45
Operator : JC/MD
Sample : L4227-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-02-102820

Quant Time: Oct 30 07:43:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

