

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102920\
 Data File : VN064376.D
 Acq On : 29 Oct 2020 16:09
 Operator : JC/MD
 Sample : PB132688TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132688TB

Manual Integrations
 APPROVED

MMDadoda
 10/30/2020 1:35:33 PM

Quant Time: Oct 30 07:45:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	236382	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	418492	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	188955	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
35) Dibromofluoromethane	7.55	113	138951	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	10.06	98	532004	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	12.38	95	234755	58.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.66%	
Target Compounds						
20) Methylene Chloride	4.50	84	8585m	3.187	ug/l	Qvalue
43) Isopropyl Acetate	8.13	43	88496	12.135	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102920\
Data File : VN064376.D
Acq On : 29 Oct 2020 16:09
Operator : JC/MD
Sample : PB132688TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132688TB

Manual Integrations
APPROVED

MMDadoda
10/30/2020 1:35:33 PM

Quant Time: Oct 30 07:45:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

