

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102920\
 Data File : VN064380.D
 Acq On : 29 Oct 2020 17:44
 Operator : JC/MD
 Sample : PB132688ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132688ZHE#11

Manual Integrations
 APPROVED

MMDadoda
 10/30/2020 1:35:39 PM

Quant Time: Oct 30 07:56:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	236598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	413542	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	427651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	174990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	188427	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	7.55	113	134884	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	10.06	98	533009	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
62) 4-Bromofluorobenzene	12.38	95	229514	57.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.42%	
Target Compounds						
20) Methylene Chloride	4.51	84	6856m	2.543	ug/l	
43) Isopropyl Acetate	8.13	43	87513	12.144	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102920\
Data File : VN064380.D
Acq On : 29 Oct 2020 17:44
Operator : JC/MD
Sample : PB132688ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132688ZHE#11

Manual Integrations
APPROVED

MMDadoda
10/30/2020 1:35:39 PM

Quant Time: Oct 30 07:56:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

