

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069291.D
 Acq On : 29 Oct 2021 10:47
 Operator : JC/MD
 Sample : VN1029MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1029MBL01

Quant Time: Oct 30 05:38:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	313726	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	552465	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	529728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	201362	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	212176	46.851	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.700%
35) Dibromofluoromethane	8.024	113	162306	47.129	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.260%
50) Toluene-d8	10.443	98	686788	49.877	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.760%
62) 4-Bromofluorobenzene	12.733	95	247899	47.709	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069291.D
Acq On : 29 Oct 2021 10:47
Operator : JC/MD
Sample : VN1029MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1029MBL01

Quant Time: Oct 30 05:38:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

