

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069299.D
 Acq On : 29 Oct 2021 14:21
 Operator : JC/MD
 Sample : M4390-01 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FRAC-COMP

Quant Time: Oct 30 05:41:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	296899	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	524707	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	519446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	195706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	202366	47.217	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.440%
35) Dibromofluoromethane	8.024	113	153534	46.940	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.880%
50) Toluene-d8	10.443	98	656727	50.217	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.440%
62) 4-Bromofluorobenzene	12.733	95	238722	48.374	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.740%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069299.D
Acq On : 29 Oct 2021 14:21
Operator : JC/MD
Sample : M4390-01 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FRAC-COMP

Quant Time: Oct 30 05:41:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

