

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069304.D
 Acq On : 29 Oct 2021 16:30
 Operator : JC/MD
 Sample : PB140327TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140327TB

Quant Time: Oct 30 05:43:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	287990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	519317	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	519445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	181003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	198278	47.694	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.380%
35) Dibromofluoromethane	8.024	113	149265	46.109	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.220%
50) Toluene-d8	10.443	98	655613	50.652	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.300%
62) 4-Bromofluorobenzene	12.733	95	233500	47.807	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.620%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	24066	13.587	ug/l	100
20) Methylene Chloride	5.095	84	15118	4.450	ug/l	97
43) Isopropyl Acetate	8.566	43	30307	3.834	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069304.D
Acq On : 29 Oct 2021 16:30
Operator : JC/MD
Sample : PB140327TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140327TB

Quant Time: Oct 30 05:43:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

