

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069305.D
 Acq On : 29 Oct 2021 16:56
 Operator : JC/MD
 Sample : PB140327ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140327ZHE#01

Quant Time: Oct 30 05:43:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	290992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	525390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	544760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	208806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	201824	48.047	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.100%
35) Dibromofluoromethane	8.024	113	151750	46.334	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.660%
50) Toluene-d8	10.443	98	668312	51.036	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.080%
62) 4-Bromofluorobenzene	12.733	95	252239	51.046	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.100%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	24401	13.634	ug/l	97
20) Methylene Chloride	5.098	84	14796	4.310	ug/l	93
43) Isopropyl Acetate	8.566	43	32489	4.035	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069305.D
Acq On : 29 Oct 2021 16:56
Operator : JC/MD
Sample : PB140327ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140327ZHE#01

Quant Time: Oct 30 05:43:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

