

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069308.D
 Acq On : 29 Oct 2021 18:13
 Operator : JC/MD
 Sample : PB140327ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140327ZHE#04

Quant Time: Oct 30 05:45:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	315317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	573123	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	574274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	221043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	216711	47.611	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.220%
35) Dibromofluoromethane	8.024	113	162976	45.618	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.240%
50) Toluene-d8	10.443	98	718600	50.306	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.620%
62) 4-Bromofluorobenzene	12.734	95	261732	48.556	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.120%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	23518	12.127	ug/l	98
18) Methyl Acetate	4.873	43	9813	1.489	ug/l	96
20) Methylene Chloride	5.095	84	15302	4.114	ug/l	97
43) Isopropyl Acetate	8.571	43	34308	3.921	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069308.D
Acq On : 29 Oct 2021 18:13
Operator : JC/MD
Sample : PB140327ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140327ZHE#04

Quant Time: Oct 30 05:45:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

