

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069312.D
 Acq On : 29 Oct 2021 19:56
 Operator : JC/MD
 Sample : M4343-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 METRO-SP

Quant Time: Oct 30 05:46:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	291672	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	527734	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	488406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	213314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	201826	47.935	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.880%
35) Dibromofluoromethane	8.018	113	146633	44.573	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.140%
50) Toluene-d8	10.440	98	642637	48.858	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.720%
62) 4-Bromofluorobenzene	12.733	95	235396	47.426	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.860%
Target Compounds						
18) Methyl Acetate	4.862	43	22566	3.702	ug/l	Qvalue # 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069312.D
Acq On : 29 Oct 2021 19:56
Operator : JC/MD
Sample : M4343-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
METRO-SP

Quant Time: Oct 30 05:46:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

