

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
 Data File : VN069313.D
 Acq On : 29 Oct 2021 20:21
 Operator : JC/MD
 Sample : M4348-09
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 370

Quant Time: Oct 30 05:46:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	305048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	546236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	511433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	219546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	210326	47.763	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.520%
35) Dibromofluoromethane	8.019	113	148756	43.687	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	87.380%
50) Toluene-d8	10.441	98	674145	49.517	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.040%
62) 4-Bromofluorobenzene	12.734	95	241500	47.008	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.020%
Target Compounds						
18) Methyl Acetate	4.865	43	16765	2.630	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102921\
Data File : VN069313.D
Acq On : 29 Oct 2021 20:21
Operator : JC/MD
Sample : M4348-09
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
370

Quant Time: Oct 30 05:46:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

